



TEST REPORT

Sample Details		Customer Details	
Sample No.	AW/000413	Company name	M/s.Manav Mandirs Smt N R P Sheth
ULR No	TC1461626000000703F		Multipurpose High School.
Sample Name	Drinking Water	Address	Balaji Heritage Chsl Sector-11 Khargar
Date of Sample Receipt	14-02-2026	Packing Condition	Sealed App 500 ml in plastic bottle
Date(s) of Performance	14-02-2026 To 16-02-2026	Other Information	NA
Date of Report Issue	17-02-2026		

Test Results:


AS PER : IS:10500-2012

Sr. No.	Test Parameters	Units	Test Methods	Limits (IS 10500 : 2012)	Results
Chemical					
1	Colour	Hazen Units	IS 3025 (Part 4)	Max 5.0	1
2	Odour	-	IS 3025 (Part 5)	Unobjectionable	Unobjectionable
3	Turbidity	NTU	IS 3025 (Part 10)	Max 1.0	<1.0
4	pH @ 25°C	-	IS 3025 Part 11	Min 6.5 - Max 8.5	7.68
5	Total Hardness (as CaCO ₃)	mg/l	IS:3025 (Part 21)	Max 200.0	30.30
6	Iron (Fe)	mg/l	IS 3025 Part 53	Max 1.0	<0.1
7	Chlorides (as Cl)	mg/l	IS 3025 Part 32	Max 250.0	8.34
8	Total Dissolved Solids (TDS) @ 180°C	mg/l	IS 3025 Part 16	Max 500.0	76
9	Calcium (as Ca)	mg/l	IS:3025 (Part 40)	Max 75.0	8.02
10	Sulphate (as SO ₄)	mg/l	IS:3025 (Part 24) sec 1	Max 200.0	1.09
11	Magnesium (as Mg)	mg/l	IS:3025 (Part 46)	Max 30.0	2.44
12	Total Alkalinity (as CaCO ₃)	mg/l	IS 3025 Part 23	Max 200.0	27.38
13	Electrical Conductivity	us/cm	IS:3025 (Part 14)	-----	142

Remark: THE SAMPLE CONFORM WITH RESPECT TO THE SPECIFICATION AND IS FIT FOR HUMAN CONSUMPTION.

Reviewed by

Pranita Parab - QA/QC Manager



Authorised Signatory



Mr. M Kapadia - Managing Director

*A. Sinha***ANURADHA SINHA**

Principal

Manav Mandir's Smt. N.R.P. Sheth**Multipurpose High School**

Affiliation No. 1130493

School Code : 30429

ATLAS/QR/ 64.4 Iss. Dt.22.08.2025

The results relate only to the items / samples tested. Sample will be retained for 30 days from the date of sample submission. The report should not be reproduced except in full without the written approval of the laboratory. The Laboratory is following the latest version of the latest method.

B - 515, 5th Floor, Mahesh Commercial Complex, Sector 15,
Next to Croma, CBD Belapur, Navi Mumbai 400 614. INDIA.
T : +91-93214 24026 / +91-85911 86267 E: contact@atlaslab.in
W: www.atlaslab.in

This Certificate is issued by the Company is subject to the General Conditions of Services stipulated overleaf. This Inspection and / or Testing reflects our findings at the time and place of intervention only and does not relieve the parties from their contractual responsibilities.



ATLASpec Labs LLP

Nationally & Internationally Approved Testing Laboratory

TEST REPORT



Sample Details		Customer Details	
Sample No.	AW/000413	Company name	M/s.Manav Mandirs Smt N R P Sheth Multipurpose High School.
Sample Name	Drinking Water	Address	Balaji Heritage Chsl Sector-11 Khargar
Date of Sample Receipt	14-02-2026	Packing Condition	Sealed App 500 ml in plastic bottle
Date(s) of Performance	14-02-2026 To 16-02-2026	Other Information	NA
Date of Report Issue	17-02-2026		

Test Results:


AS PER : IS:10500-2012

Sr. No.	Test Parameters	Units	Test Methods	Limits (IS 10500 : 2012)	Results
Chemical					
1	Appearance	-	Visual	-----	Clear Colourless Water With Small Particles
2	Taste	-	IS 3025 Part 7& 8	Agreeable	Agreeable
Microbiological Tests					
3	E.Coli	cfu/100 ml	IS 15185	Absent	Absent
4	Total Coliforms	cfu/100 ml	IS 15185	Absent	Absent

Remark: THE SAMPLE CONFORM WITH RESPECT TO THE SPECIFICATION AND IS FIT FOR HUMAN CONSUMPTION.

Reviewed by

Pranita Parab - QA/QC Manager



Authorised Signatory

Mr M Kapadia - Managing Director



A. Sinha

ANURADHA SINHA

Principal

**Manav Mandir's Smt. N.R.P. Sheth
Multipurpose High School**

Affiliation No. 1130493

School Code : 30429

The results relate only to the items / samples tested. Sample will be retained for 30 days from the date of sample submission. The report should not be reproduced except in full without the written approval of the laboratory. The Laboratory is following the latest version of the latest method.

B - 515, 5th Floor, Mahesh Commercial Complex, Sector 15,
Next to Croma, CBD Belapur, Navi Mumbai 400 614. INDIA.
T : +91-93214 24026 / +91-85911 86267 E: contact@atlaslab.in
W: www.atlaslab.in

This Certificate is issued by the Company is subject to the General Conditions of Services stipulated overleaf. This Inspection and / or Testing reflects our findings at the time and place of intervention only and does not relieve the parties from their contractual responsibilities.

End of Report

Page No. : 1 of 1

BRIHANMUMBAI MUNICIPAL CORPORATION

Solid Waste Management Department (D ward)

ACD/AE/12-23/SWM/D Dated. 20/03/2026.

SANITARY CONDITION CERTIFICATE

It is certified that, an inspection team headed by Shri. Shekhar Khedekar (Junior Supervisor SWM 'D' Ward) from Brihanmumbai Municipal Corporation inspected Manav Mandir Smt. N R P Sheth Multipurpose High School on dated. 18.03.2026 and found Manav Mandir Smt. N R P Sheth Multipurpose High School is maintaining hygienic sanitation condition, in the school building and the campus as per norms prescribed by BMC.

The school building has 18 WC's , 20 Urinals and 17 Wash Basins. They are cleaned minimum four times in a day – morning, after short and long break and after school and additionally as many times as when required. Every toilet has sanitizers and hand wash. Girls and female staff toilets have incinerators for disposing sanitary pad.

The above certificate is valid for a period of 01 year.

Signature with seal


A.E. (SWM) 'D' WARD 20/03/26

Name : Shri Amit Gadekar

Designation – Assitant Engineer (SWM) D Ward

Name and address of office –

BMC 'D' Ward SWM Department, Ground Floor

Nana Chowk, Grant Road, Mumbai - 400007



ANURADHA SINHA

Principal

**Manav Mandir's Smt. N.R.P. Shet:
Multipurpose High School**

Affiliation No. 1130493

School Code : 30429

BRIHANMUMBAI MUNICIPAL CORPORATION

Solid Waste Management Department (D ward)

ACD/AE/12-23/SWM/D Dated. 20/03/2026.

SANITORY CONDITION CERTIFICATE

It is certified that, an inspection team headed by Shri. Shekhar Khedekar (Junior Supervisor SWM 'D' Ward) from Brihanmumbai Municipal Corporation inspected Manav Mandir Smt. N R P Sheth Multipurpose High School on dated. 18.03.2026 and found Manav Mandir Smt. N R P Sheth Multipurpose High School is maintaining hygienic sanitation condition, in the school building and the campus as per norms prescribed by BMC.

The school building has 18 WC's , 20 Urinals and 17 Wash Basins. They are cleaned minimum four times in a day – morning, after short and long break and after school and additionally as many times as when required. Every toilet has sanitizers and hand wash. Girls and female staff toilets have incinerators for disposing sanitary pad.

The above certificate is valid for a period of 01 year.

Signature with seal


A.E. (SWM) 'D' WARD 20/03/26

Name : Shri Amit Gadekar

Designation – Assitant Engineer (SWM) D Ward

Name and address of office –

BMC 'D' Ward SWM Department, Ground Floor

Nana Chowk, Grant Road, Mumbai - 400007



ANURADHA SINHA

Principal

**Manav Mandir's Smt. N.R.P. Shet:
Multipurpose High School**

Affiliation No. 1130493

School Code : 30429